

Disposable Medical Protective Coverall

一、 Product name, model and specification

Product name: Disposable Medical Protective Coverall

Model: conjoined type (sterile), conjoined type (non-sterile), split type (sterile), split type (non-sterile), conjoined type with foot cover (sterile), conjoined type with foot cover (non-sterile), split type with foot cover (sterile), split type with foot cover (non-sterile)

Specification: S/160、M/165、L/170、XL/175、XXL/180、XXXL/185、XXXXL/190

二、 Name, address, production address, contact information and after-sales service unit of the registered person/manufacturer

- 1、 Name: Hunan Zhenheyikang Medical Instrument Co.,LTD
- 2、 Address: No.6 Building Jingxiang Energy, No.55 Xianguang east road, Gaoxin district Xiangtan, Hunan
- 3、 Production address: NO.6 Building, NO.8 building, 3rd and 4th floor Jingxiang Energy, No.55 Xianguang east road, Gaoxin district Xiangtan, Hunan, NO 19 Chutian Road, Yuetang District, Xiangtan City, Hunan
- 4、 Contact number: +86 17711627084
- 5、 After-sales service unit: Hunan Zhenheyikang Medical Instrument Co.,LTD

三、 Manufacturing License NO: Hunan Medical Products Administration Production Licence No. 20150013

四、 Product Registration Certificate No, Product Technical Requirements No: HNFDA Certified 20202140095

五、 Structure and composition, and the scope of application

The product is made of breathable membrane non-woven fabric with a thickness of more than 50 g / square meter, which is processed by suture, heating or bonding. The structure is composed of hooded jacket, trousers and foot covers, and the cuffs and feet are treated with tight mouth.

Ethylene oxide sterilization, sterile supply.

For barrier and protection for patients with potentially infectious blood, body fluids and secretions that are exposed to by clinical medical personnel at work.

六、 Use method: open the bag can be used

七、 Contraindications, precautions, warnings and reminders

Product contraindications: None

Damaged packaging prohibited use, expired products prohibited use;

Disposable use, please destroy after use

Avoid high temperature, humidity, direct sunlight

八、 Interpretation of graphics, symbols, abbreviations and other contents used in medical device labels



一次性使用



参考使用说明书



包装破损切勿使用



环氧乙烷灭菌



无菌

九、 Installation and use instructions or illustrations

Protective clothing does not have flame-retardant properties. It is recommended to use no more than 8 hours.

Instructions for Use: 1. Check the validity period and integrity of the protective clothing, select the appropriate specifications; 2. Read the instructions and precautions before wearing; 3. Open the protective clothing, pull the zipper from the top and down to loosen it and pull the zipper to the appropriate position; 4. Hold the toes, put the feet into the protective clothing legs; 5. Pull the protective clothing, put the arms into the sleeves of the protective clothing, do not zip and wear the cap; 6. wear other facial protective equipment

(completely cover the protective equipment); 7. Put on a protective clothing hat, zip up from bottom to top, and stick to the zipper door. Pull the zipper and seal the zipper mouth; 8. Check the suitability of protective clothing.

1. Remove the protective clothing by untying the tape and zip the cap up to remove the hat from the head; 2. Take off the sleeves and roll the protective clothing into the top from the top, facing the inside of the protective clothing; 3. Throw the removed protective clothing into the medical waste container.

十、 Product maintenance and maintenance methods, special storage, transportation conditions, methods

Products should be stored in relative humidity $\leq 80\%$, non-corrosive gas and well-ventilated indoor, ordinary transport, in the transportation, storage process should be kept clean and pollution-free, strictly prohibited rain and snow immersion.

十一、 Date of production, shelf life or expiration date

Valid for two years, production batch number, production date, expiry date please refer to the small package or qualification label.

十二、 Accessories list, including accessories, appendix, lossitage replacement cycle and replacement method description

This product has no accessories

十三、 Date and version of preparation or revision of the specification

This specification was revised in February 2023, version 2023.